Dr. Mitsching - Laboratory for Hygiene & Microbiology

Examinations - Consulting - Expert Opinions

Dr. Mitsching, Ernst-Thälmann-Str. 89, 99423 Weimar

GETT Gerätetechnik GmbH z. H. Anette Gläser Mittlerer Ring 1 08233 Treuen

2018-01-23

Test report nr.: 2017/19676

Test sample: InduProof® Smart Classic

TKG-106-IP68-WHITE-USB-DE (item no. KG19268)

Description of test product: white computer keyboard with 106 buttons, coated with silicone,

completely waterproof

Manufacturer: GETT Gerätetechnik GmbH

Analyse request: The keyboards are considered for use in fields with high hygienic

requirements, therefore it was examined if they are practicable for disinfection after application of germs and high organic load.

Date of delivery: 2017-08-02

Storage of test product: at room temperature as packed by the manufacturer

Start of testing: 2017-11-08 End of testing: 2017-11-12

Testing method:

The testing procedure was based on the germ carrier test for surface disinfection without mechanical action according to method nr 14.1 of the "Requirements and Methods for VAH certification of Chemical Disinfection Procedures" (2015-04-02).

VAH = Verbund für Angewandte Hygiene (Alliance for Applied Hygiene, former German Society for Hygiene and Microbiology)

Procedure of testing:

1) Contamination of the keyboard with test germs

In each case contamination was carried out on a single keyboard button. A gram-positive and a gram-negative bacterial test strain as well as a yeast with clinical relevance were choosen for testing.

The following test strains were used:

Staphylococcus aureus
Pseudomonas aeruginosa
DSM 799 = ATCC 6538 (= gram-positive bacterium)
DSM 939 = ATCC 15442 (= gram-negative bacterium)

- Candida albicans DSM 1386 = ATCC 10231 (= yeast)

<u>Test report nr.: 2017/19676</u> page 2 of 4

Preparation of the germ suspensions:

The bacteria suspensions were adjusted to a germ content of 1.5 - 5.0 x 10⁹ cfu / ml, whereas the suspension of Candida albicans was adjusted to 1.5 - 5.0 x 10⁸ cfu / ml.

1 ml of the organic load solution (containing 3 % bovine serum albumine and 3 % sheep erythroctes) were added to 9 ml of each germ suspension. This means an organic load of 0.3 % bovine serum albumine and 0.3 % sheep erythroctes in the test suspension (= high organic load). For each test germ 0.05 ml of this test suspensions were applied on four keyboard buttons of the same size.

- 2) After surface drying of the test suspension the contaminated spots were covered with 0.1 ml of the particular disinfectant. In parallel one contaminated spot per germ was treated with water of standardised hardness (WSH = positive control).
- 3) After an exposure time of five minutes, the survival of germs on the tested area was examined. Therefore the test spots were wiped off with swabs. Each swab was transferred into 5 ml of a media containing disinfectant neutralising reagents (3 % Tween 80 + 0.3 % lecithine + 0.1 % histidine + 0.5 % sodium thiosulphate). The germ content in this solution was examined after a neutralisation time of five minutes.

The following three disinfectants based on different active ingredients were chosen:

Active ingredient	disinfectant	manufacturer	Application concentration
			of disinfectant
alcohol (37.0 % n-propanol +	Destix MA61	Kleinmann	undiluted
24.0 % ethanol)		GmbH	
quaternary ammonium com-	Wofasept FL	Kesla	7.5 %
pounds (10 % quats)	•	Hygiene AG	
alkaline activated peracetic acid	1+1 Wofasteril	Kesla	System of two compo-
component A: 11 - 15 % peracetic	SC super	Hygiene AG	nents;
acid			0.5 % of each component
component B: alkali			

Test results:

test strain: Staphylococcus aureus

disinfectant	exposure time	Reisolated germs		Germ reduction
(active ingredient)		cfu/test spot	lg cfu/test spot	lg cfu
water of standardised hardness (positive control)	5 min	8.55 x 10 ⁷	7.93	-
Destix MA61 (alcohol)	5 min	5	0.70	7.23
Wofasept FL (quats)	5 min	325	2.51	5.42
1+1 Wofasteril SC super (peracetic a-cid)	5 min	30	1.48	6.45

cfu = colony forming units

Tel.: +49 (0) 3643 778283 Fax: +49 (0) 3643 778284 Email: labor-dr.mitsching@t-online.de Homepage: www.labor-dr-mitsching.de

<u>Test report nr.: 2017/19676</u> page 3 of 4

test strain: Pseudomonas aeruginosa

disinfectant	exposure time	Reisolated germs		Germ reduction
(active ingredient)		cfu/test spot	lg cfu/test spot	lg cfu
water of standardised hardness (positive control)	5 min	3.80 x 10 ⁷	7.58	-
Destix MA61 (alcohol)	5 min	< 5	< 0.70	> 6.88
Wofasept FL (quats)	5 min	1.45 x 10⁴	4.16	3.42
1+1 Wofasteril SC super (peracetic a-cid)	5 min	< 5	< 0.70	> 6.88

test strain: Candida albicans

disinfectant	exposure time	Reisolated germs		Germ reduction
(active ingredient)		cfu/test spot	lg cfu/test spot	lg cfu
water of standardised hardness (positive control)	5 min	1.90 x 10 ⁶	6.28	-
Destix MA61 (alcohol)	5 min	< 5	< 0.70	> 5.58
Wofasept FL (quats)	5 min	215	2.33	3.95
1+1 Wofasteril SC super (peracetic a- cid)	5 min	< 5	< 0.70	> 5.58

cfu = colony forming units

Evaluation of test results:

According to the guidelines for germ carrier tests under practical conditions of the VAH a disinfectant is considered sufficiently effective if it yields to a 5-lg reduction for bacteria and a 4-lg reduction for *Candida albicans*.

Normally tests with spray desinfectants (surface disinfection without mechanical action) are performed with plain metal germ carriers. This test had to be slightly modified for the examination of the keyboard.

The required germ reduction for bacteria and yeast could be fulfilled with the disinfectants based on alcohol and peracetic acid for all three test strains within an exposure time of five minutes.

Tel.: +49 (0) 3643 778283 Fax: +49 (0) 3643 778284 Email: labor-dr.mitsching@t-online.de Homepage: www.labor-dr-mitsching.de

<u>Test report nr.: 2017/19676</u> page 4 of 4

The testing with the disinfectant *Wofasept FL* (based on quaternary ammonium compounds) did not succeed sufficiently into the required reduction of *Pseudomonas aeruginosa* and *Candida albicans* within an exposure time of five minutes.

Summary of test results:

The keyboard TKG-106-IP68-WHITE-USB-DE (item no. KG19268) is fully coated with silicone which is practicable for cleaning and disinfection.

Due to the watercoated surface the keyboard can be submerged completeley into cleaning and disinfectant solutions. A disinfection can also be achieved by spraying the disinfectant onto the surface of the keyboard.

The intention of the testing was to examine how well the silicone surface can be decontaminated by common active ingredients in disinfectants under practical conditions.

Therefore the surface of the keyboard was contaminated with a suspension of the different test strains and organic load consisting of bovine serum albumine and sheep erythrocytes.

Without previous cleaning the buttons with the surface dried suspensions were covered with a layer of the chosen test disinfectants.

After an exposure time of five minutes the survival of germs on the tested area was examined and calculated in relation to the control surfaces treated with water of standardised hardness (WSH).

All three tested desinfectants led to an obvious germ reduction compared to the positive control with WSH.

However, taking the VAH's required germ reduction rates as basis, only the products based on alcohol as well as peracetic acid reached an adequate disinfection within five minutes exposure time.

A sufficient effect of the disinfectant based on quats could not be achieved under the tested conditions. Maybe the requested germ reduction rates can be achieved by an extension of exposure time for these active ingredients. This has to be highly considered due to the fact that quaternary ammonium compounds (quats) are often used as disinfectants in medical practice and hospitals.

In summary a good to very good practicability for disinfection can be attested to the examined keyboard. Hence, the keyboard ist suitable for use in hygienically sensible areas (e. g. medical, pharmaceutical and food industry).

Disinfectants based on alcohol or peracetic acid are recommended for this keyboard. In doing so, disinfectants listed by VAH should be used. An audition of the testing procedure under practical conditions (e. g. efficiency control with contact plates) is recommended.

Dr. M. Mitsching - head of laboratory -

The test results in this test report only relate to the tested samples and the examined parameters during the test period.

Tel.: +49 (0) 3643 778283 Email: labor-dr.mitsching@t-online.de Fax: +49 (0) 3643 778284 Homepage: www.labor-dr-mitsching.de